

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK



0324

Accredited to  
ISO/IEC 17025:2005

### Transmille Ltd

Issue No: 030 Issue date: 20 March 2013

Unit 4  
Select Business Centre  
Lodge Road  
Staplehurst  
Kent  
TN12 0QW

Contact: Mr M A Bailey  
Tel: +44 (0)1580 890700  
Fax: +44 (0)1580 890711  
E-Mail: sales@transmille.com  
Website: www.transmille.com

Calibration performed at the above address only

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
<b>ELECTRICAL</b>			
DC Resistance			
Specific Values	1 m $\Omega$ 10 m $\Omega$ 100 m $\Omega$ 1 $\Omega$ 10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ 10 M $\Omega$ 100 M $\Omega$ 1 G $\Omega$ 10 G $\Omega$ 100 G $\Omega$	280 n $\Omega$ 1.0 $\mu\Omega$ 1.7 $\mu\Omega$ 3.0 $\mu\Omega$ 28 $\mu\Omega$ 200 $\mu\Omega$ 1.5 m $\Omega$ 16 m $\Omega$ 350 m $\Omega$ 6.1 $\Omega$ 50 $\Omega$ 3.4 k $\Omega$ 220 k $\Omega$ 45 M $\Omega$ 510 M $\Omega$	100 V 100 V
Other Values	0 $\Omega$ to 10 $\Omega$ 10 $\Omega$ to 100 $\Omega$ 100 $\Omega$ to 1 k $\Omega$ 1.0 k $\Omega$ to 10 k $\Omega$ 10. k $\Omega$ to 100 k $\Omega$ 100 k $\Omega$ to 1 M $\Omega$ 1.0 M $\Omega$ to 10 M $\Omega$ 10 M $\Omega$ to 100 M $\Omega$ 100 M $\Omega$ to 1 G $\Omega$	22 ppm + 84 $\mu\Omega$ 19 ppm + 680 $\mu\Omega$ 12 ppm + 1.4 m $\Omega$ 12 ppm + 13 m $\Omega$ 14 ppm + 120 m $\Omega$ 19 ppm + 3.5 $\Omega$ 69 ppm + 165 $\Omega$ 580 ppm + 4.2 k $\Omega$ 0.58 % + 100 k $\Omega$	
AC Resistance	40 Hz to 1.592 kHz		
Specific Values	1 $\Omega$ 10 $\Omega$ 100 $\Omega$ 1 k $\Omega$ 10 k $\Omega$	41 $\mu\Omega$ 410 $\mu\Omega$ 4.1 m $\Omega$ 41 m $\Omega$ 410 m $\Omega$	



0324  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Transmille Ltd**  
**Issue No: 30 Issue date: 20 March 2013**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
DC Voltage			
Standard cell values	1.018 V nominal	1.4 $\mu$ V	This uncertainty can be realised with cells only if they have their own temperature-controlled enclosure maintained at a nominal 30 °C with the appropriate thermal stability
Other Values	0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1.0 V to 1 kV	440 nV 490 nV 2.0 ppm + 440 nV 3.0 ppm	
High Voltage	1 kV to 20 kV 20 kV to 40 kV	57 V 110 V	
DC Current	0 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 1 mA 1.0 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1.0 A to 10 A 10. A to 30 A	6 ppm + 130 pA 5.9 ppm + 1.8 nA 5.9 ppm + 26 nA 5.9 ppm + 185 nA 8.8 ppm + 1.6 $\mu$ A 36 ppm + 27 $\mu$ A 130 ppm + 500 $\mu$ A	
	10 A to 1000 A	0.23 % + 1.3 A	For the calibration of clamp-on ammeters
AC Voltage	40 Hz to 1 kHz 1 mV to 999 mV	230 ppm + 4.0 $\mu$ V	Generation only
Specific Values	1 V 10 Hz 40 Hz 1 kHz 10 kHz 100 kHz 1 MHz	40 $\mu$ V 40 $\mu$ V 40 $\mu$ V 40 $\mu$ V 230 $\mu$ V 660 $\mu$ V	
	10 V 10 Hz 40 Hz 1 kHz 10 kHz 100 kHz 200 kHz 1 MHz	390 $\mu$ V 390 $\mu$ V 100 $\mu$ V 410 $\mu$ V 2.3 mV 4.2 mV 4.1 mV	
	20 V 40 Hz 1 kHz 10 kHz 100 kHz	780 $\mu$ V 780 $\mu$ V 800 $\mu$ V 4.6 mV	



0324  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Transmille Ltd**  
**Issue No: 30 Issue date: 20 March 2013**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC Voltage (continued)			
Specific values (continued)	100 V 40 Hz 1 kHz 10 kHz 50 kHz	6.5 mV 6.4 mV 6.4 mV 78 mV	
	200 V 40 Hz 1 kHz 10 kHz 50 kHz	13 mV 13 mV 13 mV 160 mV	
	1000 V 40 Hz 1 kHz 10 kHz 50 kHz	95 mV 95 mV 95 mV 200 mV	
Other values	10 mV to 100 mV 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 20 kHz	200 ppm + 7.0 $\mu$ V 290 ppm + 5.0 $\mu$ V 250 ppm + 5.0 $\mu$ V	
	100 mV to 1 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz	80 ppm + 48 $\mu$ V 85 ppm + 25 $\mu$ V 160 ppm + 25 $\mu$ V 950 ppm + 20 $\mu$ V 1.2 % + 100 $\mu$ V	
	1 V to 10 V 10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 100 kHz 100 kHz to 1 MHz	85 ppm + 450 $\mu$ V 88 ppm + 220 $\mu$ V 170 ppm + 220 $\mu$ V 940 ppm + 300 $\mu$ V 1.2 % + 1.0 mV	
	10 V to 100 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz	230 ppm + 2.3 mV 230 ppm + 2.3 mV 870 ppm + 2.0 mV	
	100 V to 1 kV 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz	470 ppm + 20 mV 700 ppm + 25 mV 0.14 % + 20 mV	
	1 kV to 28 kV 50 Hz	0.42 % + 4.0 V	



0324  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Transmille Ltd**  
**Issue No: 30 Issue date: 20 March 2013**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AC Current	<i>40 Hz to 1 kHz</i> 25 $\mu$ A to 100 $\mu$ A 100 $\mu$ A to 1 mA 1.0 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1.0 A to 10 A 10 A to 30 A  <i>50 Hz</i> 10 A to 1000 A	150 ppm + 4 nA 160 ppm + 30 nA 160 ppm + 300 nA 160 ppm + 3.0 $\mu$ A 170 ppm + 20 $\mu$ A 350 ppm + 1.0 mA 360 ppm + 1.0 mA  0.25 % + 1.3 A	For the calibration of clamp-on ammeters
Loop impedance	<i>At 50 Hz:</i> 0.6 $\Omega$ to 1.6 $\Omega$ 5.5 $\Omega$ to 100 $\Omega$ 1 k $\Omega$	24 m $\Omega$ 42 m $\Omega$ 5.8 m $\Omega$	Nominal values for the calibration of earth loop testers
Inductance			
Specific Values	<i>1 kHz</i> 10 $\mu$ H 100 $\mu$ H 1 mH 10 mH 100 mH 1 H	8.9 nH 39 nH 350 nH 3.7 $\mu$ H 31 $\mu$ H 280 $\mu$ H	Specific values are those that fall within 1 % of the stated values.
Capacitance			
Specific Values, three-terminal	<i>1 kHz</i> 10 pF 100 pF 1 nF	26 fF 170 fF 280 fF	Specific values are those that fall within 1 % of the stated values
Specific Values, two- and three-terminal	<i>1 kHz</i> 10 nF 100 nF 1 $\mu$ F	2.8 pF 28 pF 280 pF	
Other Values	<i>1 kHz</i> 10 pF to 10 $\mu$ F	0.050 % + 0.20 pF	
Frequency			
Measurement	10 mHz to 1 GHz	0.24 ppm	The CMC is for an average frequency measured or generated over a 10-minute period. The uncertainties may be increased for shorter periods.
Generation	1 Hz to 10 MHz	2.1 Hz	



0324  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Transmille Ltd**  
**Issue No: 30 Issue date: 20 March 2013**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
Time Interval	20 ms to 900 ms	390 $\mu$ s	For the calibration of RCD testers
Phase Measurement Voltage : Current	40 Hz to 50 Hz -180 0° to +180°	170 m°	Voltage range: 1 V to 500 V Current range: 10 mA to 30 A
Temperature simulation			
Reference junction measurements	Reference at 0 °C Ambient 18 °C to 28 °C	0.12 °C 0.12 °C	
Thermocouple type			Including Reference Junction Compensation
B	100 °C to 1820 °C	2.4 °C	
E	0°C to 800 °C	0.35 °C	
J	-180 °C to +150 °C 150 °C to 750 °C	0.35 °C 0.50 °C	
K	-140 °C to +200 °C 200 °C to 1340 °C	0.40 °C 0.60 °C	
N	-270 °C to +260 °C 260 °C to 1300 °C	0.35 °C 0.60 °C	
R	100 °C to 1700 °C	1.2 °C	
S	50 °C to 1700 °C	1.7 °C	
T	-250 °C to +400 °C	0.40 °C	
Resistance Thermometer			
PT 100	-100 °C to +800 °C	0.020 °C	
END			



0324

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Transmille Ltd**

**Issue No: 30 Issue date: 20 March 2013**

**Calibration performed at main address only**

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

*A CMC is a calibration and measurement capability available to customers under normal conditions:*

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or*
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.*

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
- As an explicit function of the measurand or of a parameter (see below).
- As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
- As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

**Expression of CMCs - symbols and units**

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples, and an indication of how they are to be interpreted, are shown below.

DC voltage, 100 mV to 1 V:  $0.0025 \% + 5.0 \mu\text{V}$ :

Over the range 100 mV to 1 V, the CMC is  $0.0025 \% \cdot V + 5.0 \mu\text{V}$ , where  $V$  is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa:  $0.0036 \% + 0.12 \text{ ppm/MPa} + 4.0 \text{ Pa}$

Over the range 0.5 MPa to 140 MPa, the CMC is  $0.0036 \% \cdot p + (0.12 \cdot 10^{-6} \cdot p \cdot 10^6) + 4.0 \text{ Pa}$ , where  $p$  is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means  $1.5 \cdot 0.01 \cdot i$ , where  $i$  is the instrument indication.